

Development of automatic analysis apparatus for triple oxygen isotopes of dissolved oxygen

*Masanori Ito¹, Hiroki Sakuma¹, Daisuke Komatsu², Fumiko Nakagawa¹, Urumu Tsunogai¹, Toyoho Ishimura³

1. Graduate School of Environmental Studies, Nagoya University, 2. Department of Marine and Earth Science, School of Marine Science and Technology, Tokai University, 3. National Institute of Technology, Ibaraki College

Oxygen molecules (O_2) consists of triple oxygen isotopes (mass numbers 16, 17 and 18) providing additional unique information such as triple oxygen isotopic compositions ($\Delta^{17}O = \ln(\delta^{17}O + 1) - 0.518\ln(\delta^{18}O + 1)$). In most of the terrestrial processes (e.g. photosynthesis and respiration) fractionate O isotopes in a mass-dependent way, such that ^{17}O enrichment is about half of the ^{18}O enrichment relative to ^{16}O . As a result, $\delta^{17}O$ and $\delta^{18}O$ in terrestrial materials plot along a single line with a mass-dependent slope of about 0.52. In contrast to these mass-dependent processes, ultraviolet-induced interactions among O_2 , O_3 , and CO_2 in the stratosphere cause mass-independent fractionation with equal lowering of $\delta^{17}O$ and $\delta^{18}O$ in atmospheric O_2 . Therefore, $\Delta^{17}O$ of photosynthetically-produced O_2 in the hydrosphere shows higher values of about +150 - 250 per meg compared to atmospheric O_2 . Since the $\delta^{17}O$ and $\delta^{18}O$ of O_2 fractionated by respiration vary along a line with a mass-dependent slope, which means the $\Delta^{17}O$ will not change, we can estimate a mixing ratio of O_2 produced from photosynthesis in the hydrosphere ($\Delta^{17}O = \text{ca. } +150 \sim 250 \text{ per meg}$) and atmospheric O_2 ($\Delta^{17}O = \text{ca. } +150 \sim 250 \text{ per meg}$) dissolved in water. This will make it possible to estimate gross primary production in the lake and ocean or the air-water gas exchange coefficient by measuring the $\Delta^{17}O$ of dissolved O_2 . In this study, we constructed the new purge and trap system to measure $\Delta^{17}O$ of dissolved O_2 . The system is fully automated, extracting dissolved gases from the water samples, separate O_2 from all the other gases including Ar, and collecting pure O_2 using a cryogenic temperature cooling sampling device (ca. 10K). We will report $\Delta^{17}O$ values of dissolved O_2 in Lake Biwa where remarkable eutrophication and hypoxia have been observed in recent years.

Keywords: dissolved oxygen, triple oxygen isotopes, vertical profile