Preliminary gravity monitoring by using the superconducting gravimeter iGrav without the helium refrigeration system

*後藤 宏樹¹、池田 博²、杉原 光彦¹ *Hiroki Goto¹, Hiroshi Ikeda², Mituhiko Sugihara¹

1. 産業技術総合研究所、2. 筑波大学

1. National Institute of Advanced Industrial Science and Technology, 2. University of Tsukuba

Microgravity monitoring is a valuable tool for mapping the redistribution of subsurface mass associated with geothermal exploitation, hydrocarbon production, volcanic activity, groundwater movement, and other subsurface processes. The superconducting gravimeter is currently the most precise gravimeter and thus is a useful instrument for microgravity monitoring. In this study, we tested a method of measuring gravity changes by using the superconducting gravimeter iGrav manufactured by GWR Instruments from which the helium refrigeration system (cryogenic refrigerator and compressor) was removed to make it more portable. The iGrav was installed in the Geological Survey of Japan, National Institute of Advanced Industrial Science and Technology (AIST), Tsukuba. Gravity monitoring was conducted for 6 days. Helium level in the dewar after the measurement was 40 %. The measured gravity changes were almost consistent with the calculated gravity changes by the GOTIC2 program (Matsumoto et al., 2001). Two weights were placed at a distance of 10 and 40 cm from the upper end of the iGrav to check whether or not the measured gravity changes caused by known changes in mass distribution were explained quantitatively. The measured gravity change (0.9 μ Gal) was consistent with the calculated one (0.8 - 1.0 μ Gal). For further studies, applying the method of measuring gravity changes by using the iGrav without the helium refrigeration system to a field-scale study is required.

References:

Matsumoto, K., Sato, T., Takanezawa, T., and Ooe, M. (2001), GOTIC2: A program for computation of oceanic tidal loading effect, Journal of the Geodetic Society of Japan, 47, 243-248.

キーワード:超伝導重力計、重力モニタリング、ヘリウム冷却システム Keywords: Superconducting gravimeter, Gravity monitoring, Helium refrigeration system